

CORRIGENDUM

Date-02-08-2018

With reference to the Minutes of Meeting held on 01-08-2018 at GVK EMRI GOA head office, related to tender no-GVKEMRI/GOA/03/SCM/2018-19 dated on 27-07-18, named as "**Short Term Tender notice for Procurement of medical equipment & other**".

The following modifications/clarifications may be take note of prior to submission of bids by the prospective bidders.

Sr.No.	References	Query raised by vendor	Tender committee decision
1	Page no.16, Defibrillator , pt.no.1,	Requested by vendor to consider European CE certified equipment's also.	Committee agreed to accept either European CE or US FDA certified equipment's
2	Page no.16, Defibrillator, pt.no.17 ,	Clarification related to the requirement of ETCO2	Provision is mandatory for future up gradation
3	Page.no.16, Defibrillator , pt.no.3,	Requested to change weight of the Defibrillator Should not be more than 10Kgs with all accessories	Committee agreed to accept the change request. Defibrillator with less than 10 kgs will be accepted
4	Page.no.16, Defibrillator , pt.no.9,d)	Requested to erase -Touch screen desirable specification	Touch screen is Desirable but not mandatory
5	Page.no.17, Defibrillator , pt.no.11,a), ii)	Query for minimum 100mm for comfort of viewing /examination to be considered.	Higher than 48-55 mm will be considered
6	Page.no.17, Defibrillator , pt.no.16 , i)	Query for European CE with four digit notified code to be considered.	European CE certification will be considered
7	Page.no.17, Defibrillator, pt.no.16 , j)	Query related to charging time should be less 7 seconds	No change
8	Page.no.18, Defibrillator , pt.no.18 , g)	Query related to E-1789 which is the certified mount for mounting in Ambulance to be considered.	Mount not required as it is in ECC. Can be on crash cart. Optional
9	Page.no.17, Defibrillator, pt.no.19	Query related to possibility in case of manufacturing defect only	Replacement in case of faulty equipment during warranty period
10	Page.no.17, Defibrillator , pt.no.20	Query related to possibility in case of manufacturing defect only	Replacement in case of faulty equipment during warranty period

Note: Bidder need to submit the corrigendum document along with tender document with signature and stamp as token of acceptance of modifications/ clarifications.

Sd/

GVK EMRI-GOA