

CORRIGENDUM

Date-22-05-2026

With reference to the Minutes of Meeting held on 20-05-2026, 2.30pm at EMRI GHS -Goa , related to tender no- Tender No: EMRI GREEN HEALTH SERVICES /GOA/01/SCM/2026-27 Dated: 17.05.2026, named as “Short Term Tender notice for procurement of Medical Equipments & others , Goa “

The following modifications/clarifications may be take note of prior to submission of bids by the prospective bidders.

Sr.No.	Reference	Query raised by vendor	Tender committee decision
1	Page no. 4, 11)	Requested the quoted rates shall remain valid for maximum period of 30 days from the date of submission / opening	Its mandatory as per the requirement
2	Page no.16, 10),c)	Request to have CMS central monitoring system via 5 G	It will be additional optional specification , apart from declared minimum specification
3	Page no.17, 16),h)	Request to have without US FDA	US FDA or EU Certification is mandatory
4	Page no.17, 16),4)	Request to have 110mm printer as it is required to print 12 led ECG	Its mandatory as per the requirement
5	Page no.18, 2),3)	Requested to have weight should be 2kg onwards	Its mandatory as per the requirement
6	Page no.18, 2),5)	Requested to have demand, CMV, SMMV are also activated if the patient breathes spontaneously	It will be additional optional specification , apart from declared minimum specification
7	Page no.18, 2),8)	Requested to have Lithium 3.6V battery pack for alarm 7 technical indication purpose	Its mandatory as per the requirement
8	Page no.18, 3),a)	Requested to have ISO - 13485/CE/European CE / CDSCO/BIS certified	Its mandatory as per the requirement
9	Page no.18, 3),c)	Requested to have + / - 1.5% or more accuracy	Its mandatory as per the requirement
10	Page no.18, 3),f)	Requested to have AC mode only	Its mandatory as per the requirement
11	Page no.18, 3),g)	Requested to have battery back up more than 6 hrs as using pump for transportation purpose	Its mandatory as per the requirement
12	Page no.1 , 1)	on behalf of Director of health services(DHS), Goa	Request to exclude this sentence
13	Page no.1 , 5),c.	Pre bid conference at DHS	Read as " Pre bid conference at EMRI GHS -Goa

Note: Bidder need to submit the corrigendum document along with tender document with signature and stamp as token of acceptance of modifications/ clarifications.

Sd/

**EMRI Green Health
Services -GOA**